

Multifunction instrument for safety, functionality and performance verifications on a PV plant

The multifunction instrument PVCHECK performs prompt and safe electrical checks required for a PV system (DC section) and controls of the functionality of modules / strings in accordance with IEC/EN62446 guideline

PVCHECK: safety checks

PVCHECK verifies the continuity of the protective conductors (and associated connections) and measures the insulation resistance of the active conductors on a module, a string, or a photovoltaic field in accordance to IEC/EN62446 guideline, without the need of any external switch to short-circuit the positive and negative terminals.



PVCHECK: functionality checks

PVCHECK verifies the functionality of a PV string in accordance to the EN62446 guideline by measuring the open circuit voltage and the short-circuit current at operating conditions and extrapolating the results to the STC (by measuring the solar radiation). Finally, it displays the measurements and a comparison to the PV strings previously tested.



PVCHECK: performance checks

PVCHECK analyses the performance of a PV array (DC) under the operating conditions (connected to the inverter) displaying the generated power and the efficiency of the PV plant in accordance to the IEC/EN62446.



PVCHECK performs safety checks, functionality checks and performance checks on a PV plant



With the remote unit SOLAR-02 the irradiance and module/environment temperature measured values are shown also in “independent mode” (ideal solution during a pre-test on the installation) beside the test/recording with PVCHECK

The HT304N reference cell allows to perform solar irradiance measurements both on Monocrystalline and Polycrystalline PV modules



2. ELECTRICAL SPECIFICATIONS

Accuracy is calculated as $\pm [\% \text{ readings} + (\text{no. of digits}) * \text{resolution}]$ at $23^{\circ}\text{C} \pm 5^{\circ}\text{C}$, relative humidity $<80\% \text{HR}$

2.1. PERFORMANCE TEST

DC Voltage

Range (V)	Resolution (V)	Uncertainty
15.0 ÷ 999.9	0.1	$\pm (0.5\% \text{rdg} + 2 \text{dgt})$

DC current (by mean external clamp)

Range (mV)	Resolution (mV)	Uncertainty
-1100 ÷ -5	0.1	$\pm (0.5\% \text{rdg} + 0.6 \text{mV})$
5 ÷ 1100		

DC current is always positive ;DC current zeroed if the related voltage value is $< 5 \text{mV}$

DC Power ($V_{\text{meas}} > 150 \text{V}$)

Clamp FS (A)	Range (W)	Resolution (W)	Uncertainty
$1 < \text{FS} \leq 10$	0.000k ÷ 9.999k	0.001k	$\pm (0.7\% \text{rdg} + 3 \text{dgt})$ ($I_{\text{meas}} < 10\% \text{FS}$)
	10.00k ÷ 99.99k	0.01k	
$10 < \text{FS} \leq 100$	0.000k ÷ 9.999k	0.001k	$\pm (0.7\% \text{rdg})$ ($I_{\text{meas}} \geq 10\% \text{FS}$)
	10.00k ÷ 99.99k	0.01k	
$100 < \text{FS} \leq 1000$	0.00k ÷ 99.99k	0.01k	
	100.0k ÷ 999.9k	0.1k	

Irradiance (by mean HT304N - Sensitivity = k)

Range (mV)	Resolution (mV)	Uncertainty
0 ÷ 16.0 ($k < 10$)	0.01	$\pm (1.0\% \text{rdg} + 0.1 \text{mV})$
0 ÷ 50.0 ($k \geq 10$)	0.02	

Temperature (by mean PT300N)

Range ($^{\circ}\text{C}$)	Resolution ($^{\circ}\text{C}$)	Uncertainty
-20.0 ÷ 100.0	0.1	$\pm (1.0\% \text{rdg} + 1^{\circ}\text{C})$

2.2. FUNCTIONALITY TEST

DC Voltage @ OPC

Range (V)	Resolution (V)	Uncertainty (*)
5.0 ÷ 199.9	0.1	±(1.0%rdg+2dgt)
200 ÷ 999	1	

(*) Measure starts for VDC > 15V, Uncertainty defined for VDC > 20V

DC Current @ OPC

Range (A)	Resolution (A)	Uncertainty
0.10 ÷ 10.00	0.01	±(1.0%rdg+2dgt)

DC Voltage @ STC

Range (V)	Resolution (V)	Uncertainty (*)
5.0 ÷ 199.9	0.1	±(4.0%rdg+2dgt)
200 ÷ 999	1	

(*) Measure starts for VDC > 15V, Uncertainty defined for VDC > 20V

DC Current @ STC

Range (A)	Resolution (A)	Uncertainty
0.10 ÷ 10.00	0.01	±(4.0%rdg+2dgt)

Irradiance (by mean HT304N - Sensitivity = k)

Range (mV)	Resolution (mV)	Uncertainty
0 ÷ 16.0 (k < 10)	0.01	±(1.0%rdg + 0.1mV)
0 ÷ 50.0 (k ≥ 10)	0.02	

Temperature (by mean PT300N)

Range (°C)	Resolution (°C)	Uncertainty
-20.0 ÷ 100.0	0.1	± (1.0%rdg + 1°C)

2.3. SAFETY TEST

Continuity Test (LOWΩ)

Range [Ω]	Resolution [Ω]	Uncertainty
0.00 ÷ 1.99	0.01	±(2.0%rdg + 2dgt)
2.0 ÷ 19.9	0.1	
20 ÷ 199	1	

Test current >200mA DC up to 2Ω (test leads included), Resolution 1mA, Uncertainty ±(5.0%rdg + 5dgt)

Open loop voltage 4 < V_o < 24V

Insulation Test (MΩ)

Test voltage [V]	Range [MΩ]	Resolution [MΩ]	Uncertainty
250	0.01 ÷ 1.99	0.01	±(5.0%rdg + 5dgt)
	2.0 ÷ 19.9	0.1	
	20 ÷ 99	1	
500	0.01 ÷ 1.99	0.01	
	2.0 ÷ 19.9	0.1	
	20 ÷ 199	1	
1000	0.01 ÷ 1.99	0.01	
	2.0 ÷ 19.9	0.1	
	20 ÷ 199	1	

Generated voltage

Resolution 1V, Uncertainty ±(5.0%rdg + 5dgt) @ R_{mis} > 0.5% FS

Test current

> 1mA with load = 1kΩ x V_{nom}



3. GENERAL SPECIFICATIONS

DISPLAY AND MEMORY:

Features: 128x128pxl custom LCD with backlight
Memory: max 999 test

POWER SUPPLY:

PV CHECK internal power supply: 6x1.5V alkaline batteries type LR6, AA, AM3, MN 1500
Number of Test : > 999 test
SOLAR-02 power supply: 4x1.5V alkaline batteries type AAA LR03
SOLAR-02 max recording time (@ IP=5s): approx. 1.5h

OUTPUT INTERFACE

PC communication port: optical/USB
Interface with SOLAR-02: wireless RF communication (max distance 1m)

MECHANICAL FEATURES

Size (L x W x H): 235 x 165 x 75mm
Weight (batteries included): 1.2kg

ENVIRONMENTAL CONDITIONS:

Reference temperature: $23^{\circ}\text{C} \pm 5^{\circ}\text{C}$
Working temperature: $0^{\circ} \div 40^{\circ}\text{C}$
Working humidity: <80%HR
Storage temperature (remove the batteries): $-10 \div 60^{\circ}\text{C}$
Storage humidity: <80%HR

GENERAL REFERENCE STANDARDS:

Safety: IEC/EN61010-1
Safety of measurement accessories: IEC/EN61010-031
Measurements: IEC/EN62446 (PV performance, IVCK)
IEC/EN 61557-1, 2, -4 (LOW Ω , M Ω)
Insulation: double insulation
Pollution degree: 2
Overvoltage category: CAT I 1000V DC, CAT III 300V to ground
Max 1000V among inputs P, N, E, C
Max altitude of use: 2000m

This instrument complies with the requirements of the European Low Voltage Directives 2006/95/EEC (LVD) and EMC 2004/108/EEC